Operation and Maintenance of the EPV100 Portable Ventilator

Applications of the EPV100

- The EPV100 Portable Ventilator is a gas powered electronically controlled mechanical ventilator, designed to provide emergency respiratory support by means of a face mask or an endotracheal tube inserted into a patient's airway.
- The EPV100 is a volume controlled ventilator that delivers timedcycled constant flow breaths.
- The EPV100 is an all-weather ventilator suitable for use at the scene of a medical incident, as well as in pre-hospital, intra-hospital, and inter-hospital transport.
- The EPV100 has been designed to withstand direct exposure to rain, up to 100G shock and vibration, and drops of less than 4 feet.
- These ventilators are intended for use on patients weighing greater than 20kg, or 44 lbs.

Overview of Controls and Settings



General Specifications

Tidal Volume: 200 to 1200 ml

Breaths Per Minute: 8 to 30 (1-second inspiratory time)

or 8 to 20 (2-second inspiratory time)

Unit Run Time: Run time with settings of 640 ml

D Cylinder
Jumbo D Cylinder
E Cylinder
Hospital O,

Approximately 64 minutes
Approximately 99 minutes
Approximately 105 minutes
Unlimited run time

Hospital O₂ Unlimited run time Hospital Air Unlimited run time

Manometer Readout: 5 to 99 cm H₂O

Size: 3.5" deep x 7.0 wide x 9.3" high

(88.9 x 177.8 x 236.2 mm)

Weight: 3.1 lbs (1.4 kg)

Inspiratory Time: 1 second or 2 seconds

Battery Run Time: 48 hours at room temperature with settings

of 10 BPM and 2-second inspiratory time;

2 D batteries required

NOTE: The EPV100 is powered by compressed gas, using 2 D batteries to control the alarms and settings.

3

Notes on Controls and Settings

- The EPV100 is powered by compressed gas but requires 2D cell batteries to power the controls and settings. Therefore it must have both batteries as well as a compressed oxygen source to run.
- It features electronic power management alarms to provide advance notice of low gas and low battery status.
- The EPV100 has independent control of Tidal Volume, Inspiratory Time, and Breaths Per Minute.
- It features electronic ventilation alarms to provide notice of dangerous or ineffective ventilation.
- The EPV100 does not have an Assist Control function, but will allow the patient to take spontaneous breaths while connected to the patient circuit.
- The unit also features a digital manometer to monitor the peak airway pressure during each breath.



EPV100 Connections

Ventilator Connections

Oxygen Connection



- A 40-80 psi compressed oxygen source <u>must</u> be connected to the EPV100 for operation of this ventilator.
- Connect a compressed oxygen source by attaching one end of an oxygen hose to the DISS oxygen inlet shown above, and then connecting the other end to an oxygen regulator, or hospital or ambulance oxygen outlet.
- Ensure that the oxygen source is pressurized and capable of delivering oxygen at 40-80 psi.

Ventilator Connections

Patient Circuit Connection



- To attach the breathing circuit, press the open end of a three-foot circuit with one-way valve firmly on to the 22 mm patient circuit connection.
- The circuit connection will accommodate any standard 22 mm ventilation circuit with a one-way valve, although the EPV100 has been tested and approved using the Allied three-foot breathing circuits with one-way valve.

Ventilator Connections


Patient Circuits

- Allied offers several affordable three-foot circuit configurations to meet the caregiver's requirements.
- All circuits contain three feet of corrugated tubing, a one-way duck-bill valve, an expiratory port capable of accepting a PEEP valve or bacterial filter, and a swivel connector capable of accepting a mask or ET tube.
- Optional configurations incorporate a pre-packaged cuffed mask and/or a bacterial exhalation filter.
- The following part numbers represent the various pre-packaged configurations and accessory items offered.

three-foot Ci	rcuit Configurations for EPV100 Portable Ventilator	
Allied Part #		Case Quantity
L599-140	3 foot circuit, One Way Valve, Swivel Connnector, Adult cuffed Mask, Exhalation Filter	10
L599-190	3 foot circuit, One Way Valve, Swivel Connnector, Exhalation Filter (No Mask)	10
L599-130	3 foot circuit, One Way Valve, Swivel Connnector, Adult cuffed Mask (No Filter)	10
L599-180	3 foot circuit, One Way Valve, Swivel Connnector (No Mask or Filter)	10
Circuit Acces Allied Part #	sories Available From Aliied Healthcare Products, Inc.	Case Quantity
LPEEP	Adjustable PEEP Valve, 0-20cm H ₂ O	12
L599-200	599-200 Exhalation filter	
L595161-10	Adult Oxygen Mask	10
L595162-10	Child Oxygen Mask	10

Patient Circuit Connections

Overview of Allied Patient Circuit Connections




Patient Circuit Connections

Fitting of PEEP Valve and Mask or ET Tube

- An External PEEP valve may be fitted to the expiratory port of the patient circuit, allowing adjustable PEEP from 0-20 cm H₂O.
- The swivel connector on the three-foot Allied patient circuit features a 22 mm universal adapter that will accept a standard cuffed mask or endotracheal tube.
- To fit the mask or ET tube, press the end of the mask or tube into the open end of the swivel connector.
- The patient circuit also contains a removable access port. Ensure that this port is closed during ventilation to prevent loss of airway pressure.

Battery Installation





- For the EPV100 to function, two D cell batteries must be installed in the battery compartment. Two D cell batteries are supplied with the unit.
- The batteries will provide 48 hours of run time under ventilation parameters of 10 BPM and 2-second Inspiratory Time, and 640 ml Tidal Volume.
- To install the batteries, loosen the four brass thumb screws on the battery compartment cover.
- Remove the cover and install the batteries as indicated with the positive terminals facing toward the top of the unit as shown.
- Replace the cover with the operating instructions facing up, and tighten the thumb screws to create a positive seal of the rubber gasket.

11

Start-Up

Powering On the Unit




- Ensure that a 40-80 psi gas source is attached and pressurized, and the batteries are installed.
- Ensure that a patient circuit is installed (to prevent alarms from activating, the circuit output must be attached to a test lung or device that provides similar compliance to human lungs).
- Press the power button to power on the unit.
- The unit will go through a start-up mode in which all the LEDs and lights will flash, and then the unit will begin to operate according to the set parameters.

EPV100 Controls and Settings

Controls and Settings


Inspiratory Time Adjustment



- The Inspiratory Time is selectable at 1 second or 2 seconds, and color coded for ventilation of Child (Orange) or Adult (Black) patients.
- To change the Inspiratory Time setting, press the desired 1 SEC or 2 SEC button on the keypad.
- A green LED will indicate the selected inspiratory time

Controls and Settings

Breaths Per Minute Adjustment




- Adjustable from 8-20 Breaths Per Minute when inspiratory time is set to 2 Sec.
- Adjustable from 8-30 Breaths Per Minute when inspiratory time is set to 1 Sec.
- To change the Breaths Per Minute Setting, press the up or down arrows on the keypad near the BPM label.
- Recommended AHA guidelines for ventilation:

Adult: 8-12 BPM

Child: 12-20 BPM

Controls and Settings

Tidal Volume Adjustment




- The Tidal Volume is adjustable from 200-1200 ml and is color keyed for 1 Second or 2 Second ventilation
- If your inspiratory time is set to 2 seconds (Adult), refer to the Black numerical markings and select the most appropriate tidal volume from the choices on the dial.
- If the inspiratory time is set to 1 second (Child), refer to the Orange numerical markings and select the most appropriate tidal volume from the choices on the dial.
- The following is a table of approximate settings for tidal volume by patient height, derived from ideal body weight with oxygen delivery at 10ml/Kg.


This table is an approximation for reference only. Refer to direction of a physician or medical professional for appropriate settings.

Height Adult Male (in feet' and Inches")	4' 8"	4' 11"	5' 3"	5' 6"	5' 10"	6' 1"	6'5"	6' 8"	6' 11"	7' 3"	7' 6"
Height Adult Female (in feet' and Inches")	4' 10"	5' 1"	5' 5"	5' 8"	6' 0"	6' 5"	6' 8"	6' 10"			
Approximate Tidal Volume Setting	400	480	560	640	720	800	880	960	1040	1120	1200

Height Child Male (in feet' and Inches")	3' 11"	4' 1"	4' 2"	4' 4"	4' 6"	4' 8"	4' 9"	4' 11"	5' 1"
Height Child Female (in feet' and Inches")	4' 1"	4' 3"	4' 4"	4' 6"	4' 8"	4' 10"	4' 11"	5' 1"	5' 3"
Approximate Tidal Volume Setting	200	240	280	320	360	400	440	480	520

- The EPV100 features the following alarms:
 - High airway pressure
 - Low airway pressure
 - Low battery (visual alarm only)
 - Low Source Gas
- All alarms on the EPV100 are pre-set and non-adjustable.
- Alarms are signaled audibly and visibly except for the low battery alarm.
- When an alarm is activated a 60 Db audible alarm will sound, and the appropriate red alarm LED will illuminate, signaling the type of alarm activated.
- Audible alarms can be silenced for 110 seconds by pressing and holding the selected inspiratory time button for 3 seconds (the one that has a green LED).





60 Db Alarm Tone

Flashing Alarm Status LEDs


- High Airway Pressure Alarm (audible/visual):
 - Activates at 45 cm H₂O, indicating that the patient airway is obstructed or lungs have becoming non-compliant. Check the connections and airway.
 - Will sound a consistent tone and the High Airway Pressure LED will illuminate
 - Note that the EPV100 features a safety pressure relief that will limit airway pressures to less than 60 cm H₂O, to prevent damage to the lungs.
 - This alarm is automatically cleared when 25 seconds pass without a high airway pressure being detected.




- Low Airway Pressure Alarm (audible/visual):
 - Activates if the airway pressure drops below 9 cm H_2O for >15 seconds
 - This could indicate a circuit disconnect or an open access port in the circuit. Check the connections.
 - Will sound a consistent tone and the High Airway Pressure LED will <u>flash</u>.
 - Note that the digital manometer readout will also display the low airway pressure.
 - The alarm will clear when airway pressure exceeds 9 cm H_2O .



- Low Source Gas Alarm (audible/visual):
 - Activates between 40 psi and 35 psi (275 to 241kPa) source pressure, indicating that the gas source is critically low and supplemental compressed gas should be sought immediately. The run time remaining depends on the size of the cylinder used, but in most cases is less than one minute remaining.
 - It will sound a consistent tone and the Low Source Gas LED will illuminate
 - The alarm will clear when source gas pressure is restored to greater than
 40 psi.
 - Should the source gas fall below 35 psi, the unit may cease to function.




- Low Battery Alarm (visual only):
 - Activates when there are less than 2 hours of battery life remaining.
 - There is no audible alarm; however, the Low Battery LED will illuminate indicating that the batteries should be replaced.



Spontaneous Breathing

• The EPV100 does not have an assist control function, but will allow the patient to take spontaneous breaths while connected to the patient circuit via an anti-suffocation valve located at the base of the patient circuit connection.



- It is important that the anti-suffocation valve remains unobstructed during ventilation.
- The resistance to spontaneous breathing through the antisuffocation valve is minimal. The following table indicates the inspiratory resistance at different flow rate demands.

Inspiratory Resistance During a Spontaneous Breath

Demanded inspiratory flow rate (LPM)	15	30	45	60
Inspiratory Resistance (cm H ₂ O)	0.6	1.7	3.3	5.1

Power Management

EPV100 Run Times

- The EPV100 run time is dependent on the amount of source gas present and limited by the battery life.
- Two D cell batteries at full capacity will run the EPV100 continuously for 48 hours at 10 BPM, 2 second Inspiratory Time and 640 ml Tidal Volume. The batteries are standard Alkaline type and replacement of the batteries will provide an additional 48 hours of run time.
- If the batteries are depleted to a point in which the unit cannot provide effective ventilation, the EPV100 will close the delivery solenoid to stop gas flow, and stop delivering breaths, resulting in an alarm.
- This shut-down sequence will be preceded by redundant visible/audible alarms to provide advance notice to locate and replenish the supplies.

Power Management

EPV100 Run Times (continued)

- The EPV100 does not dump oxygen during normal operation, therefore the run time of the unit from the source gas is directly related to the minute volume of oxygen per the Tv, It, and BPM selections.
- As a reference, the following table indicates the approximate run times for one EPV100 at 10 BPM, and 640 ml tidal volume.

Cylinder Size	Capacity (Liters)	Approximate Run Time
D Cylinder	425	65 minutes
Jumbo D Cylinder	647	99 minutes (1.5 hours)
E Cylinder	684	105 minutes (1.75 hours)
M Cylinder (ambulance)	3454	531 minutes (8.75 hours)
		48 hours before battery
Hospital Oxygen Outlet	Virtually unlimited	replacement is needed

Initial Operation Procedure

Initial Operation and Check-Out Procedure

The unit should be checked for proper operation upon receipt as well as before each use. The following procedure can be performed upon receipt and after cleaning to prepare the unit for the next use.

- Install the batteries.
- Set the ventilator to the following settings:
 - -BPM = 10
 - Tidal Volume = 640
 - Inspiratory Time = 2 seconds
- Connect a 50 psi oxygen source to the unit, pressurize the hose, and turn the power on; it should begin to cycle.
- After 20 seconds confirm that the low airway pressure alarm has sounded and the low airway pressure LED is flashing.

Initial Operation and Check-Out Procedure

- Let the vent cycle for 15 seconds.
- Confirm that the breath delivery alarm LED is flashing and an audible alarm has sounded, indicating that the low airway pressure alarm has tripped.
 - Press and hold the 2 sec Inspiratory Time button for 3 seconds to silence the alarm for 110 seconds.
- Using a stop watch, count the number of breaths delivered in one minute (60 seconds).
 - Confirm that between 9 and 11 breaths have been delivered.
 - Confirm that the inspiratory time is significantly shorter than the expiratory time. (At the settings noted above, the ventilator should provide a 2.0 second inspiratory time and a 4.0 second expiratory time)

Initial Operation and Check-Out

continued

- Occlude the patient circuit connection port with the palm of your hand.
 - Confirm that the high airway pressure alarm LED is illuminated and an audible alarm has sounded.
 - Confirm that the airway pressure does not exceed 60 cm $\rm H_2O$ by monitoring the airway pressure display on the face of the EPV100 ventilator. The airway pressure should read approximately 45 cm $\rm H_2O$.
 - Press and hold the 2sec Inspiratory Time button for 4 seconds to silence the alarm for 110 sec.
- Turn the source gas off.
 - Confirm that the Low Source Gas LED display is illuminated and an audible alarm has sounded.
 - Press and hold the 2 sec Inspiratory Time button for 4 seconds to silence the alarm for 110 seconds.

Initial Operation and Check-Out

 Press the power button to turn the power off, remove the oxygen hose from the inlet, and store the unit for future use.

Should the unit fail any of these tests contact Allied Healthcare Products, Inc., Technical Support Center at 1-800-411-5136 for assistance.

Routine Maintenance

Routine Maintenance

Cleaning and Disinfecting

- The EPV100 ventilator should be cleaned and disinfected after each use.
 - Wipe the unit down with a damp rag containing a mild cleaning solution to remove any residue from the surface.
 - Once the residue has been removed, the unit should be wiped with isopropyl alcohol or a cold disinfecting solution to kill bacteria.
 - The unit should then be wiped down with water to remove any film left by the cold disinfecting solution.
 - Make sure the unit is dry before putting the unit away.
- The following is a list of tested disinfecting solutions:

• Isopropyl Alcohol: 70% IPA

Alconox: I tablespoon Alconox to 1 gallon H₂O

Cetylcide: 2 tablespoons Cetylcide to 1 gallon H₂O

• Bleach: 10% bleach in H₂O

Routine Maintenance

- The EPV100 should be checked annually to ensure proper function using the calibration procedure noted in the instruction manual.
- Batteries should be checked regularly and replaced if they are beyond the expiration date on the battery. Replace with standard D cell alkaline batteries.
- After 5 years, or if annual calibration fails, the EPV100 should be serviced by an authorized repair technician.

For technical assistance or to schedule service, contact Allied Healthcare Products, Inc. Technical Support Center at 800-411-5136.

We hope this guide has prepared you to properly and safely use and maintain your EPV100 portable ventilator. If you have additional questions or require technical assistance, please contact our Technical Support Center.

Technical Support Contact Information

techsupport@alliedhpi.com

800-411-5136